

Humidity Generation and Calibration Equipment

THUNDER SCIENTIFIC®
CORPORATION THE HUMIDITY SOURCE

Model 2500

Test Chamber Fluid Filling Procedure



Tech Support Document TSD-0207
Revision Date: March 29, 2023
Rev-B

623 WYOMING BLVD. SE



ALBUQUERQUE, NM 87123-3198
WWW.THUNDERSCIENTIFIC.COM



TEL: 505.265.8701
FAX: 505.266.6203

2500 Test Chamber Fluid Filling Procedure

1. Make sure the power to the 2500 system is off.



2. Make sure any air to the system is off.




3. Remove the top right hand panel, over the chamber, using the Torx screwdriver provided.



4. Lift off the chamber top panel.



5. Locate the chamber fluid fill port.




6. Remove the fluid fill port insulation plug.



7. Remove the "Red" Plastic fill port plug cap.



8. Use a funnel that will fit into a 5/8" (15.9mm) diameter opening, or use the one provided.



9. Using 2 U.S. gallons of 70% concentrate propylene glycol fluid, pour them in and fill the remainder with “recommended” distilled water until unit level is approximately 1/4” (6.35mm) below the bottom of the fill port.



10. Turn on the power switch.




11. Turn on the air and open the air tank valve.



12. Press the RUN button on the front panel and let the system run for approximately 5-10 minutes.



13. After the system has run push the STOP button.



14. Remove the funnel and check the fluid level and see if it's at 1/4" below the fill port, top off the fluid if necessary.



15. Replace the “RED” plastic fill port plug cap.



16. Insert the fluid fill port insulation plug.



17. Replace the top right hand panel making sure the textured side is up.



18. Replace the screws using your Torx screwdriver. Chamber fluid filling is now complete.



Please Call 1-800-872-7728, Fax 1-505-266-6203, or E-mail support@thunderscientific.com should you have any questions.